YES WE HAVE MANUAL

? DATE

DO WE HAVE SCHEMATICS FOR THIS INSTR?

ELECTRODYNETRICS 518-762-2600

1) CALIBRATION NOT RIGHT - MAX IF GAIN READS ± 76 DBμV

2) WILL CHANGE 10DB WHEN IT FEELS LIKE IT TURING RF ATT OR IF GAIN KNOBS PUTS IT BACK.

3) SCHEMATIC FOR DM 1000. (ESA 1000)

4) COST FOR CAL/REPAIR $95

E
518-762-2600
231 ENTERPRISE RD
JOHNSTOWN, NY 12095

ART NOWICKI